

Issued by NMI Certin B.V.

In accordance with WELMEC 8.8 Issue 2, Paragraph 8.1 of EN 45501:1992/AC:1993, WELMEC 2.1 Issue 4, OIML R 76-1 (2006).

Producer Shanghai Caisun Electronic Co., Ltd.
No.25, No.369 Datuan Sandun Sanxuan Road
Shanghai 201312
China

Measuring instrument An **Indicator**, tested as a part of a weighing instrument.

Designation : XK315A1-2X

Further properties are described in the annexes:

- Description TC8263 revision 0;
- Documentation folder TC8263-1.

An overview of performed tests is given in the annex:

- Description TC8263 revision 0.

Issuing Authority

NMI Certin B.V.
12 March 2014



C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see "Regulation objection and appeal against decisions of NMI" www.nmi.nl)

Reproduction of the complete document only is permitted



Description

Number **TC8263** revision 0
Project number 12200515
Page 1 of 4

1 General information about the indicator

All properties of the indicator, whether mentioned or not, shall not be in conflict with the standard mentioned in the certificate.

This certificate is the positive result of the applied voluntary, modular approach, for a component of a measuring instrument, as described in WELMEC 8.8. The complete measuring system must be covered by an EC type-approval Certificate.

1.1 Essential parts

Number	Pages	Description	Remarks
8263/0-01	1	Software block diagram	-
8263/0-02	1	Hardware block diagram	-
8263/0-03	1	System interface	-
8263/0-04	3	Main board	Including parts list

EMI protection measures:

- The A/D part on the main board is shielded with a metal cover;
- Ferrite bead on the cable between main board and load cell connector.

1.2 Essential characteristics

Accuracy class	III and IIII
Maximum number of verification scale intervals	3000
Load cell excitation voltage	5 V DC
Minimum input voltage per verification scale interval	1 μ V
Minimum load cell resistance	87 Ω
Maximum load cell resistance	1220 Ω
Temperature range	-10 °C / +40 °C
Fraction of the maximum permissible error	0,5
Load cell connection	4-wire
Maximum value of the cable length per cross wire section between the indicator and the load cells	The load cells are connected directly without junction box
Weighing range(s)	Single interval
Power supply voltage	7,5 V DC supplied by an AC/DC adapter 6 V DC supplied by a battery
Software identification	Version number: VEr21.3

Software:

- The identification number will be displayed at start-up;
- The indicator has embedded software;

List of legally relevant functions:

- Determination stability of equilibrium;
- Zero indicator;
- Semi-automatic zero-setting;
- Initial zero-setting;
- Zero-tracking;
- Semi-automatic subtractive tare weighing;
- Adjustment / set-up mode via a switch on the main board;
- Acting upon significant faults;
- Checking the display;
- Extended indicating, resolution 1/10 e during pressing a key.



Description

Number **TC8263** revision 0
Project number 12200515
Page 3 of 4

1.3 Essential shapes

The indicator is built according to drawing:

Number	Pages	Description	Remarks
8263/0-05	5	Exploded view	-

The descriptive markings plate is secured against removal by sealing or will be destroyed when removed and contains at least the following information:

- This certificate number TC8263;
- Producers name or mark.

Inside the cabinet is an adjustment lock, located on the main board.

1.4 Conditional parts

The indicator may be equipped with one or more of the following protective interfaces that have not to be secured:

- RS232.

1.5 Non-essential parts

Display;
Keyboard;
Battery.

2 Seals

To secure components that may not be dismantled or adjusted by the user, the indicator has to be secured in a suitable manner on the locations indicated in the drawing:

Number	Pages	Description	Remarks
8263/0-06	5	Seals	-

The connecting cable of the load cell or the junction box is provided with possibility to seal.



Description

Number **TC8263** revision 0
Project number 12200515
Page 4 of 4

3 Conditions for conformity assessment

The compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in WELMEC 2 Issue 5 Section 11, at the time of putting into use.

Other parties may use this Evaluation Certificate only with the written permission of the producer.

4 Reports

An overview of performed tests is given in the reports:

- No. NMI-12200515-01 dated 12 March 2014 that includes 48 pages;
- No. NMI-12200515-02 dated 12 March 2014 that includes 12 pages.

A report can be a test report, an evaluation report, a type evaluation report and/or a pattern evaluation report.