

TEST CERTIFICATE

Number E-16.02.C01

LOAD CELL TYPE CS-D

Issued by: LGAI TECHNOLOGICAL CENTER S.A.
Campus de la U.A.B.
Ronda de la Font del Carme, s/n.
E-08193 BELLATERRA SPAIN
Notified Body Number 0370.

In accordance with Paragraph 3.10 of the European Standard "Metrological aspects of non-automatic weighing instruments" EN 45501:2015. The applied error fraction p_{LC} with reference to paragraph 3.10.2.1 of this standard is 0,7. Following paragraph F.2 of annex F of this standard, the tests have been performed according to the OIML International Recommendation, OIML·R 60 (2000).

Issued to: SENSOCAR S.A.
Calle Géminis, 77. P.I. Can Perellada..
E-08228 TERRASSA. ESPAÑA.

In respect of: The model of a load cell, with strain gauges, equipped with electronics, composed by modules 2, 3 and 4 in Figure 1 of the UNE-EN 45501 (april 2016), tested as part of a non-automatic weighing instrument.
Manufacturer: SENSOCAR S.A.
Type: CS-D.


Characteristics:

Classification	---	C3 ↓	C4 ↓	C5 ↓	C6 ↓	---
Maximum number of LC verification intervals	n_{LC}	3000	4000	5000	6000	---
Maximum capacity	E_{max}	10 t up to and including 50 t				---
Ratio minimum LC verification interval	$Y = E_{max}/V_{min}$	18000				---
additional marking	temperature limits	Number of counts for E_{max}	minimum dead load	safe overload		
--	-10°C/+40°C	60000	$E_{min} = 0$ kg	$E_{lim}/E_{max} = 150\%$		

The main characteristics are shown in the descriptive annex, which is an integral part of the test certificate and consists of 8 pages.

The type is described in the submitted technical documentation, identified with number 03/16 (16/34514130-D).

Managing Director
Product Conformity B.U.



Xavier Ruiz Peña
Bellaterra, 05 september 2016.

This document shall not be reproduced except in full, with the annex.
This test certificate refers only to metrological requirements.
This test certificate cannot be used without applicant's authorization.